

Compliance Details

Interaction Date: 4/14/2005 9:34 AM

Patient:

L Cflex

Patient ID 2

Home Phone
Age

Referring Physician:



Group/Practice
Phone
Fax
Email

PCP:

Phone


Clinician:

Compliance Information 4/13/2005 - 4/13/2005



Patterns of Use

April, 2005












Sleep Therapy Long Term Trend

4/13/2005 - 4/13/2005

AutoCPAP Min AutoCPAP Max AutoCPAP 90% AutoCPAP




Max AutoCPAP	15.0
Min AutoCPAP	8.0
Average Non-Responsive Apnea/Hypopnea Index	0.0
Average Obstructive Apnea Index	1.0
Average Hypopnea Index	1.0
Average AHI	1.9
Average Flow Limitation Index	1.9
Average Snore Index	1.2
Average Max Leak	89.8
Average 90% Leak	33.2
Average Leak	23.6
Average Large Leak	0 mins.





Sleep Therapy Daily Details

4/14/2005 1:12 AM - 4/14/2005 1:31 AM



Legend: NR - Non-Responsive Apnea/Hypopnea, OA - Obstructive Apnea, H - Hypopnea, FL - Flow Limitation, S - Snore

Daily Events Per Hour

4/14/2005 01:12 AM

Total AHI: 0.0

	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Pressure (cm H2O)	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Minutes at Pressure	0.0	0.0	18.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
% of Night	0.0	0.0	100.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Flow Limitation	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Snore	0.0	0.0	6.7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Non Responsive Apnea/Hypopnea	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Obstructive Apnea	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Hypopnea	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
AHI:	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0


90%





Sleep Therapy Daily Details

4/14/2005 1:41 AM - 4/14/2005 8:45 AM



Legend: NR - Non-Responsive Apnea/Hypopnea, OA - Obstructive Apnea, H - Hypopnea, FL - Flow Limitation, S - Snore

Daily Events Per Hour

4/14/2005 01:41 AM

Total AHI: 1.6

	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Pressure (cm H2O)	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Minutes at Pressure	0.0	0.0	0.0	0.0	229.5	119.5	28.5	34.5	3.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
% of Night	0.0	0.0	0.0	0.0	55.2	28.8	6.9	8.3	0.8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Flow Limitation	0.0	0.0	0.0	0.0	1.3	2.0	2.1	1.7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Snore	0.0	0.0	0.0	0.0	0.5	0.5	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Non Responsive Apnea/Hypopnea	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Obstructive Apnea	0.0	0.0	0.0	0.0	0.3	0.0	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Hypopnea	0.0	0.0	0.0	0.0	2.1	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
AHI:	0.0	0.0	0.0	0.0	2.4	0.5	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

90%



Summary of Daily Details

Daily Events Per Hour

4/13/2005 - 4/13/2005


Total AHI: 1.5

	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Pressure (cm H2O)	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20
Minutes at Pressure	0.0	0.0	18.0	0.0	229.5	119.5	28.5	34.5	3.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
% of Night	0.0	0.0	4.2	0.0	52.9	27.6	6.6	8.0	0.8	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Flow Limitation	0.0	0.0	0.0	0.0	1.3	2.0	2.1	1.7	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Snore	0.0	0.0	6.7	0.0	0.5	0.5	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Non Responsive Apnea/Hypopnea	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Obstructive Apnea	0.0	0.0	0.0	0.0	0.3	0.0	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
Hypopnea	0.0	0.0	0.0	0.0	2.1	0.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0
AHI:	0.0	0.0	0.0	0.0	2.4	0.5	2.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0

90%

Auto-CPAP Time at Pressure

4/13/2005 - 4/13/2005



Summary of Compliance - All Data

Compliance Statistics

Date Range	4/13/2005 - 4/13/2005 (1 day)
Days with Device Usage	1 day
Days without Device Usage	0 days
Percent Days with Device Usage	100 %
Cumulative Usage	7 hrs. 14 mins.
Maximum Usage (1 Day)	7 hrs. 14 mins.
Average Usage (All Days)	7 hrs. 14 mins.
Average Usage (Days Used)	7 hrs. 14 mins.
Minimum Usage (1 Day)	7 hrs. 14 mins.
Percent of Days with Usage >= 4 hours	100 %
Percent of Days with Usage < 4 hours	0 %

RemStar Auto Statistics

Auto-CPAP Mean Pressure	8.6 cm H2O
Auto-CPAP Peak Average Pressure	10.1 cm H2O
Average Device Pressure <= 90% of Time	9.9 cm H2O
Average Time in Apnea Per Day	24 secs.
Average Time in Large Leak Per Day	0 mins.